

Anexo A: Exemplos de etiquetagem UDI

Figura 1 – Exemplos de locais de rotulagem no monitor Starling e embalagens de produtos contendo o UDI incorreto (sob o código de barras):

Cabo do Paciente:



Pacote de Bateria:



Monitor Starling:



Embalagem do Monitor:



Figura 2 – Exemplo de etiqueta corrigida do Monitor Starling:

Starling Monitor		SN V512022022352
		CHEMISTS Rating: 100-240V; 50/60 Hz; 2A Cheatech Medical, Inc. 600 SE Maritime Ave Suite 220, Vancouver, WA 98661 USA Tel: (+1) 866-733-9997 Fax: (+1) 360-718-8154 EC REP: MedNet EC REP GmbH Borkstrasse 10, 48163 Münster, Germany Cheatech Medical (Israel) Ltd 2A Haplotha St, F-22, Tel-Aviv 6102020 Israel Tel: (+972) 3-646-0288 Fax: (+972) 3-725-4848 CHEETAH STARLING is Patented. Cheatech Medical, Inc. is a wholly-owned indirect subsidiary of Baxter International Inc. This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions: 1. This device may not cause harmful interference. 2. This device must accept any interference received, including interference that may cause undesired operation. Made in Israel CE mark 2017-03-01 CPN0802 Rev. 002
Baxter		

(01) 072500 13030268 (21) 512022022352 11 170301
 CTIN Serial # Mfg Date

Baxter and Starling are trademarks of Baxter International Inc.